## **CALIBRATION KIT DATASHEET**



## Calibration kit for N Series

## 6550Y09-M



Configuration	า		
Connector Type		N Male	
Connector Impedance		50 Ohms	
Connector Polarity		Standard	
Power Handling		≤2W	
Electrical Specifications			
Calibration standard	Frequency	Parameter	
Opens	DC~9GHz	Phase Deviation≤±0.8°	
Shorts	DC~9GHz	Phase Deviation≤±0.8°	
Calibration loads	DC~9GHz	Return Loss≤-36dB (VSWR≤1.032)	
Mechanical data			
Mating cycles		≥2000	
Recommended torque		1.0~1.36Nm	
Environmental Data			
Operating Temperature		+15°C ~ +35°C	
Storage temperature		-40°C ~ +75°C	

While the information has been carefully compiled to the best of our knowledge, nothing is intended as representation or warranty on our part and no statement herein shall be construed as recommendation to infringe existing patents. In the effort to improve our products, we reserve the right to make changes judged to be necessary.

Page

1/1

Gwave Technology Inc.	Tel.: +86 10 86484190
4 <sup>th</sup> Floor,New Material Plaza,Yongfeng Hi-Tech Base,Haidian	Fax: +86 10 86484190-0
District,Beijing ,China(100094) www.gwave-tech.com	email: gwave@gwave-tech.com